

**Rajiv Gandhi Institute of Petroleum Technology – Jais, Amethi**

**Corrigendum-1**

Date: 29.11.2019

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/63 in rgipt.euniwizarde.com portal & institute Website for the procurement of Vacuum Oven with Vacuum Pump.

**Corrections/Addendum to Vacuum Oven with Vacuum Pump Technical Specifications**

	<b>Before Pre-bid Meeting</b>	<b>After Pre-bid Meeting</b>
1.	<b>Tender Document Page 2 of 21</b> SI. No. 1: Oil Free Vacuum pump: Nominal pumping speed 2.0 m <sup>3</sup> /h or better, final pressure 7 mbar or better with resolution 1 mbar or better, including all necessary parts for vacuum connection, ATEX conformity: II 3G IIC T3 X	<b>Tender Document Page 2 of 21</b> SI. No. 1: Oil Free Vacuum pump: Nominal pumping speed 2.0 m <sup>3</sup> /h or better, final pressure 7 mbar or better with resolution 1 mbar or better, including all necessary parts for vacuum connection.

**N. B.: A marking System will be used to evaluate the bids. The bid that receives more than the threshold mark fixed by the purchase committee will be considered further for finance bid.**

*Shukla* 29/11/19

Dr. A. Shukla

Chairman, BSH Purchase Committee