

RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY, JAIS, AMETHI

CORRIGENDUM-1

Date: 27.06.2019

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/16 in rgipt.euniwizarde.com portal & Institute Website for the procurement of 01 No. Laboratory Freezer.

Name of Equipment	01 No. Laboratory Freezer
Submission Date	28.06.2019 (Friday) before 3pm
Date & Time of Opening of Techno- Commercial bid	28.06.2019 (Friday) at 4pm
Revised Submission Date	15.07.2019 (Monday) before 3pm
Revised Date & Time of Opening of Techno- Commercial bid	15.07.2019 (Monday) at 4pm

A. Technical Specifications for Laboratory Freezer

- 1) **Vertical Type Laboratory Freezer**
- 2) Chamber Capacity (Litres): Greater than 300
- 3) **Working Capacity (Litres): Greater than 250**
- 4) **Controller: Microprocessor with external LED graphic display**
- 5) Temperature range (°C): -40 to -10 or better
- 6) Temperature Resolution (°C): 0.1
- 7) Interior Basic: Aluminium
- 8) **Shelves: 6-8**
- 9) Max shelf workload (kg): Greater than 10
- 10) Max unit workload (kg): Greater than 50
- 11) Voltage: 230 V, 50 Hz
- 12) Nominal Power (W): 500
- 13) Additional internal glass door
- 14) Chamber: Door type solid

B. On-site Installation and Training free

C. Two years on-site warranty with spares.

Note: Highlighted portions are the changes in conditions. The other terms & conditions remain unchanged.

This issues with the approval of Competent Authority.

Sd/-

Assistant Registrar (Accounts)