

Rajiv Gandhi Institute of Petroleum Technology – Jais, Amethi

Corrigendum-1

Date: 29.11.2019

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/61 in rgipt.euniwizarde.com portal & institute Website for the procurement of Reactor.

Corrections/Addendum to Reactor Technical Specifications

	Before Pre-bid Meeting	After Pre-bid Meeting
1.	Tender Document Page 2 of 21 SI. No. 2: Max Speed: 3000 rpm	Tender Document Page 2 of 21 SI. No. 2: Max Speed: 2000 rpm or better
2.	Tender Document Page 2 of 21 SI. No. 4: Min. /Max. Temperature { -20 to +200 °C (in heating jacket)}	Tender Document Page 2 of 21 SI. No. 4: Min. /Max. Temperature { -10 to +200 °C (in heating jacket)}
3.	Tender Document Page 2 of 21 SI. No. 9: Material Seals made of Flex-O-Form (modified PTFE), Kalrez®	Tender Document Page 2 of 21 SI. No. 9: Material Seals made of Flex-O-Form, Kalrez® or FKM/FFKM
4.		Heating Circulator and temperature controller system for +20 to +200 °C with accessories.

N. B.: A marking System will be used to evaluate the bids. The bid that receives more than the threshold mark fixed by the purchase committee will be considered further for finance bid.


29/11/19

Dr. A. Shukla

Chairman, BSH Purchase Committee