

**RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY, JAIS, AMETHI**

**CORRIGENDUM-1**

**Date: 15.11.2019**

Please refer to E-Tender on RGIPT/JAIS/E-OPN/LAB/2019-20/37 in [rgipt.euniwizarde.com](http://rgipt.euniwizarde.com) portal & Institute Website for the procurement of Surface Area Measurement System.

**Corrections/Addendum to Surface Area analyser Technical Specifications**

	<b>Before Pre-bid Meeting</b>	<b>After Pre-bid Meeting</b>
1	<b>Tender Document Page 2 of 22</b> <b>SI. No. 4:</b> The system should have facility for dead volume correction without any need to maintain the liquid nitrogen level.	<b>Tender Document Page 2 of 23</b> <b>SI. No. 1.8:</b> The system should have facility for dead volume correction with appropriate technique to maintain the liquid nitrogen level.

**N.B.: A marking system will be used to evaluate the bids. The bid that receives more than the threshold mark fixed by the purchase committee will be considered further for finance bid.**

**The last date to receive the bids is extended till 28<sup>th</sup> Nov 2019 3pm & the Opening of Techno- Commercial bid will be at 4pm.**

  
Dr. U. Ojha

Chairman, Institute Equipment Purchase Committee

  
M.S. [unclear]