



**RAJIV GANDHI INSTITUTE OF PETROLEUM TECHNOLOGY JAIS
AMETHI**

CORRIGENDUM FOR THE QUANTITY OF "CEMENT TESTING EQUIPMENT"

Tender Ref No.: RGIPT/JAIS/R&D/TS/P2217/2022-23, dated 20/10/2022.


Subject: Correction in the item quantity of "Cement Testing Equipment"

This is to inform that, the quantity of items of tender for "Cement Testing Equipment" has been corrected as below in table.

Technical Specification 01 no. of Cement Testing Equipment

S.No	Specification	Quantity (number)
1	Beach Mixture <ul style="list-style-type: none">➤ 1 spindle, Speeds: 0-18,000 RPM.➤ No of Speeds: 3 (Low, Medium and High).➤ Volume: 1500 ml.➤ 220V ± 5% AC; 50 Hz; Power: 200 W.	2
2	Capillary Suction Timer <ul style="list-style-type: none">➤ Test head assembly with funnel (1 Qty).➤ Digital - readouts have higher accuracy.➤ 7- Segment or LCD Display - makes measurements easy to read.➤ Power Requirements – 9 to 24 Volt Battery or 240 V AC.	1
3	Shale Swelling Analyzer with Precision Weighing Balance <ul style="list-style-type: none">➤ Measuring Units 1 head➤ 150 ml Measuring Cylinder class 'A'➤ Holding Cell (Open) for shale Sample with Docking Clamp➤ Thermocouple➤ Calibration Block, Multi Point➤ Flat Screen; 1 1/16" Diameter➤ Teflon Washer and Transfer Stand➤ Wafer Tube and Bottom Plate➤ Cup and Cap for Wafer Tube➤ Stirring Hot Plate; 230 Volt Only➤ Magnetic Stir Bar, 1"➤ 60 cc Disposable Syringe➤ Weighing Balance for Measuring weight: repeatability 0.1 mg, range 0 to 50 gm.	1

Dr. Tushar Sharma
Chairman, Purchase Committee


31/10/2022